



## Sludge activity (TTC) cuvette test

Product #: LCK318 Hazardous

AED Price: Contact Hach







## LCK cuvette tests: Minimum amounts of reagents, maximum safety.

Pre-dosed reagents for maximum safety. Easy to use and photometric evaluation, packaged in a fully equipped box. Environmentally friendly due to less chemicals.

Precise and reliable measurement values

Maximum safety for users, thanks to the closed system and low amounts of reagents

Convenient and error-free dosing of the reagents

Barcode label for automatic recognition in the photometer

Differentiation between tests and measuring ranges by means of colour-coding

## **Specifications**

According to standard: DIN 38412-3

Description: Screening test for sludge activity (TTC-SA) or residual sludge activity (TTC-RA)

Instrument: DR3900, DR6000, DR1900, DR2800, DR3800, DR5000

Measuring range: 5 - 200 μg Formazan (SA)

Measuring range (2): 0 - 500 % Residual activity (RA)

Method: Colorimetric
Parameter: Sludge activity

Platform: LCK
Storage conditions: 2 - 8 °C

## Required Accessories

- Membrane filtration set with 50 membrane filters 1.2 μm (Item LCW904)
- Accessory kit sludge activity (Item LZC918)
- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)