



Total Chlorine DPD Reagent Powder Pillows, 10 mL, pk/1000

Product #:

2105628

Hazardous

AED Price:
Available

Contact Hach



For total chlorine determination by the DPD Method. USEPA accepted for reporting for drinking and wastewater analysis using Hach Method 8167.

Range: 0.02 - 2.00 mg/L. Sample Size: 10 mL.

Pack of 1000 powder pillows.

Boric Acid free!

Specifications

EPA compliant:	Yes
Instrument:	All except DR/3000, DR/2000
Measuring range:	0.02 - 2.00 mg/L Cl ₂
Method:	DPD
Method number:	8167
Number of tests:	1000
Parameter:	Chlorine, total
Platform:	Powder Pillows
Sample size:	10 mL
Storage conditions:	10 - 25 °C

Required Accessories

- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)