



DPD Total Chlorine Reagent Powder Pillows, 25 mL, pk/1000

Product #:

1406428

Hazardous

AED Price:
Available

Contact Hach



For total chlorine determination by the DPD method. USEPA accepted for reporting drinking and wastewater analysis using Hach Method 8167.

Ranges: Low Range: up to - 2.00 mg/L; Mid Range: 0.05 - 4.00 mg/L; High Range: 0.1- 10.0 mg/L. Sample Size: Low Range - 25 mL. Mid Range: 10 mL. High Range: 5 mL.

Pack of 1000 powder pillows.

Boric Acid free!

Specifications

EPA compliant:	Yes
Measuring range:	0.02 - 2.00 mg/L Cl ₂
Measuring range (2):	0.1 - 10.0 mg/L Cl ₂
Measuring range (3):	0.05 - 4.00 mg/L Br ₂
Method:	DPD
Method number:	8167
Method number :	10070
Number of tests:	1000
Parameter:	Total Chlorine, Bromine, Iodine
Platform:	Powder Pillows
Sample size:	25 mL
Storage conditions:	10 - 25 °C

Required Accessories

- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)