

Free Chlorine Reagent Powder Pillows, DPD (pk/1000)

Product #: **1407028** Hazardous

AED Price: Available

Contact Hach



For free chlorine determination by the DPD method.

USEPA accepted for reporting drinking water analysis using Hach Method 8021.

Ranges: LR: up to 2.00 mg/L; MR: 0.05 - 4.50 mg/L; HR: 0 .1- 10.0 mg/L.

Sample Sizes: LR: 25 mL; MR: 10 mL; HR: 5 mL.

Pack of 1000 powder pillows.

Boric Acid free!

Specifications

EPA compliant:	Yes
Measuring range:	0.02 - 2.00 mg/L Cl_2
Measuring range (2):	0.05 - 4.50 mg/L Cl_2
Method:	DPD
Method number:	8021
Method number :	10069
Number of tests:	1000
Parameter:	Chlorine, free
Platform:	Powder Pillows
Sample size:	25 mL
Storage conditions:	10 °C - 25 °C

Required Accessories

- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)