



## Free Chlorine Reagent Powder Pillows, DPD (pk/100)

**Product #:**

**1407099**

**Hazardous**

AED Price:  
Available

Contact Hach



For free chlorine determination by the DPD method. USEPA accepted for reporting drinking water analysis using Hach Method 8021.

Ranges: LR: up to 2.00 mg/L; MR: 0.05-4.00 mg/L; HR: 0.1-10.0 mg/L.

Sample Sizes: LR: 25 mL; MR: 10 mL; HR: 5 mL.

Pack of 100 powder pillows.

Boric Acid free!

---

### Specifications

Description:	DPD Free Chlorine Reagent Powder Pillows, 25mL
EPA compliant:	Yes
Footnote:	*Generally, the DR 5000 performs the same methods as the DR/4000. The DR 2800 performs the same methods as the DR 2700, DR/2500, and DR/2400. Some exceptions do exist.
Instrument:	All
Measuring range:	0.1 - 10.0 mg/L Cl <sub>2</sub>
Measuring range (2):	0.05 - 4.00 mg/L Cl <sub>2</sub>
Method:	DPD
Method number:	10245
Method number :	10069
Number of tests:	100
Parameter:	Chlorine, free
Platform:	Powder Pillows
Sample size:	25 mL
Storage conditions:	10 °C - 25 °C

---

### Required Accessories

- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)