



Manganese Reagent Set, 0.006-0.700 mg/L Mn, 25 mL

Product #:

2243300

AED Price:
Available

Contact Hach

Hazardous



For the determination of Manganese with PAN Method. For 25 mL sample.

For determination of Manganese by the PAN Method. Hach Method 8149. Range: 0.006 - 0.700 mg/L Mn. Sample Size: 25 mL. Reagent Set includes: Ascorbic Acid Powder Pillows, Alkaline-Cyanide Reagent and PAN Indicator Solution (0.1%) for 100 tests.

(DR/3000, DR/2000)

Specifications

Instrument:	DR/3000, DR/2000
Measuring range:	0.006 - 0.700 mg/L Mn
Method:	PAN
Method number:	8149
Number of tests:	100
Sample volume:	25 mL
Special Features:	Manganese Reagent set, 25ml