

Chlorine Dioxide Reagent Set, 0.04-5.00 mg/L ClO_2

Product #: **2770900** Hazardous

AED Price: Available

Contact Hach



For the determination of Chlorine Dioxide, DPD/Glycine Method. For 10mL sample.

Specifications

EPA compliant:	Yes
Instrument:	All
Measuring range:	0.04 - 5.00 mg/L ClO_2
Method:	DPD/Glycine
Method number:	10126
Number of tests:	100
Parameter:	Chlorine dioxide

Required Accessories

- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)